

Nome commerciale:

TERMOMETRO INFRAROSSI NON TOUCH

Codice Prodotto:

TER-NT



Technical parameters

Measuring range: 32.0°C~42.2°C

Accuracy: $\pm 0.2^{\circ}\text{C}$ (35.0°C~42.0°C)

$\pm 0.3^{\circ}\text{C}$ (32.0°C~34.9°C及42.1°C~42.2°C)

Resolution: 0.1°C

Memory: 32

Power supply : 2*AAA(DC 3V)

Dimension: 144mm*101mm*26mm

Weight: 88g(exclude battery)

Auto power-off: 60s

Features



a key
measure



memory
storage



replaceable
battery



a second
display



c/f
switch



forehead
measurement



hot tip



LED
display



automatic
shutdown



intelligent
accurate



Benannt durch/Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
ZLG-BS-244.10.08



Product Service

EC Certificate

Production Quality Assurance System
Directive 93/42/EEC on Medical Devices (MDD), Annex V
(Devices in Class IIa, IIb or III)

No. [REDACTED]

Manufacturer: [REDACTED]

Facility(ies): [REDACTED]

Product
Category(ies):

Clinical Infrared Thermometer,
Electronic Thermometer and Blood
Pressure Monitor

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for manufacture and final inspection of the respective devices / device categories in accordance with MDD Annex V. This quality assurance system conforms to the requirements of this Directive and is subject to periodical surveillance. For marketing of class IIb and III devices an additional Annex III certificate is mandatory. See also notes overleaf.

Report No.: [REDACTED]

Valid from:

2019-07-15

Valid until:

2022-05-07

Date,

2019-07-15

Stefan Preiß

Head of Certification/Notified Body